

Brief EU regulatory Update

Doris-Ann Williams

BIVDA

Diagnostics; making a difference



The political timetable

slide provided by the MHRA

January 2014

Greeks chair
Member State
meetings

July 2014

Italians chair
Member State
meetings

January 2015

Latvians chair
Member State
meetings

Spring 2014

European
Parliament
campaigns

May 2014


European
Parliament
elections

October 2014

New European
Commissioners

Progress!

Slide provided by the MHRA



Proposed Classification

In the new classification system, IVDs will be divided into four classes of risk:
A (lowest risk), B, C and D (highest risk)

The conformity assessment procedure for class A devices will be carried out, as a general rule, under the sole responsibility of the manufacturer

For devices of classes B, C and D an appropriate level of involvement of a Notified Body is compulsory proportionate to the risk class

Devices of class D will require explicit prior approval of the design and of the quality management system before they may be placed on the market

In the case of class B and C devices, a Notified Body is required to check the quality management system and additionally, for class C, the technical documentation of representative samples.

Clinical evidence requirements

Covered in Annexes XII and XIII

Extended from analytical validity to

Scientific validity

ie association to a clinical condition or a physiological state

(from existing tests; literature; expert opinion; results of proof of concept studies; results from clinical performance studies)

Requirement for clinical evidence report

Analytical & scientific validity + performance studies

‘Living’ document expanded and amended with post marketing data



"You're fired, Jack. The lab results just came back, and you tested positive for Coke."

Doris-Ann Williams
BIVDA

Tel 0845 618 8224

doris-ann@bivda.co.uk

www.bivda.co.uk
@BIVDA

BIVDA